

REQUEST FOR QUOTATION

For: SUPPLY AND DELIVERY OF MEDICAL AND LABORATORY EQUIPMENT FOR USE OF GEAMH

Date: October 30, 2020

RFQ No.: nlontoc-RFQ-G112-2020

PR No.: 2020-10-H-155/10-16-2020

PHILGEPS Ref. No.: 7298413

Name of Company

Complete Company Address

To whom it may concern:

Please quote your lowest price/s on the lot or item/s below, subject to General Conditions indicated herein, stating the shortest time of delivery and submit your quotation duly signed by your official representative not later than **November 05, 2020 at 5:00 PM to the BAC Office.**

Very truly yours,
RENATO A. ABUTAN
BAC-A Chairman

CANVASSERS'S CERTIFICATION

This is to certify that I have full knowledge, authority and responsibility in distributing and/or collecting the Request for Quotation in accordance to the guidelines in securing prices for **General Emilio Aguinaldo Memorial Hospital** of the Provincial Government of Cavite.

(Printed Name above Signature)
AUTHORIZED CANVASSER

PLEASE QUOTE: **PER ITEM**

BIDDER'S PROPOSAL BOX

No	ITEM DESCRIPTION (ITEM NAME&TECHNICALSPECIFICATIONS)	QTY.	UNIT	APPROVED BUDGET OF CONTRACT (ABC)	FINANCIAL PROPOSAL (Indicate the Price Offer)		TECHNICAL PROPOSAL	Delivery weeks /days
					Unit Price	Total Price		
1	SINGLE PASS MOBILE REVERSE OSMOSIS SYSTEM	1	UNIT	485,000.00				
**SEE PAGE 2 FOR CONTINUATION **								

- GENERAL CONDITIONS**
1. All entries must be legibly handwritten or typewritten;
 2. Bidders must submit necessary business permits(SEC,LGU,DTI, CDA, BIR Certificate, etc.) and Philgeps Online Registration/Certification;
 3. Place this RFQ in a sealed envelope and type the following details on the face of the envelope:
Your Company Name
RFQ No.
PR No.
PHILGEPS Reference No.
 4. Delivery period must be at least within _____ calendar days from receipt of the Notice to Proceed (Indicate the days of delivery in the Bidder's Proposal box)
 5. Item/s delivered must have warranties for unit replacements, parts, labor or other services;
 6. Price validity shall be for a period of three(3) months;
 7. Quoted prices must be inclusive of taxes, and other charges or fees and shall not exceed the Approved Budget for the Contract(ABC);
 8. If Bid is accepted, the supplier undertakes to provide a performance security in the form, amounts, and within the times specified in the Revised IRR of R.A. 9184;
 9. The Provincial Government of Cavite reserves the right to reject any and all bids, declare a failure of bidding, or not award the contract at any time prior to contract award in accordance with the provisions of RA 9184, without thereby incurring any liability to the affected bidder or bidders.
 10. Transaction with the Provincial Government of Cavite shall mean compliance by the winning bidder with the bid and delivery requirements before the issuance of check payment;
 11. Failure to comply with these conditions shall mean disqualification of your bid proposal

**SUPPLIER/CONTRACTOR/CONSULTANTS
CERTIFICATION**

Date: _____

After having carefully read and accepted Your General Conditions, I/We quote you on the item/s at price/s noted above for immediate delivery and shipment which can be made in _____ days from receipt of the Notice To Proceed.

Print Name & Signature of
Authorized Representative

Company Tel./Fax/Mobile No.

Company Tax Identification No. (TIN)

***Attach Certificate of Registration (BIR)**

304 10/29

**Page 2 of Request for Quotation for SUPPLY AND DELIVERY OF MEDICAL AND LABORATORY EQUIPMENT
FOR USE OF GEAMH**

PLEASE QUOTE: PER ITEM

BIDDER'S PROPOSAL BOX

No	ITEM DESCRIPTION (ITEM NAME&TECHNICALSPECIFICATIONS)	QTY.	UNIT	APPROVED BUDGET OF CONTRACT (ABC)	FINANCIAL PROPOSAL <i>(Indicate the Price Offer)</i>		TECHNICAL PROPOSAL	Delivery weeks /days
					Unit Price	Total Price		
	<p>Specification for Portable Reverse Osmosis System for Haemodialysis Machines</p> <p>1. Technical specifications</p> <p>1. Portable Reverse Osmosis System suitable for two machine with pure water capacity of 316 liters per hour.</p> <p>2. The unit must be 2000 gallons per day, single pass mobile reverse osmosis.</p> <p>3. Should have an automatic water control and monitoring system of reject flow, feed water and pure water conductivity (water treatment) including reverse osmosis machine.</p> <p>4. It must have an automatic control system, supplied with multi-stage vertical pump, high grade stainless steel, thin film with large spiral design and FRP pressure membrane vessel, flow meters, dial stainless oil pressure gauges, pressure regulator control valve and TDS meter.</p> <p>5. Should have following rejection ration:</p> <p>a. Hardness - 95%</p> <p>b. Bacteria and Endotoxins - 99%</p> <p>c. Pure water efficiency - 75%</p> <p>Ultra water sterilizer</p> <p>d. Total Dissolved Solid (TDS) - less than 50 ppm.</p> <p>6. Purity of treated water should be as per AAMI standard of water for hemodialysis.</p> <p>7. Stainless booster pump.</p> <p>8. Should have adequate pre-treatment stage such as:</p> <p>a. Sediment filter</p> <p>b. Carbon filter</p> <p>c. Water soft view for use with tap water</p> <p>d. Ultraviolet system/sterilizations</p> <p>9. The pre-treatment equipment must have multimedia and carbon FRP tank with manual/ automatic timer backwashing head, water softener FRP tank with manual/automatic backwashing head, water sampling and emergency bypass valve, brine tank, raw water multi-stage horizontal stainless steel pump head back-up system, and pressure regulator control valve.</p> <p>10. The whole assembly should be housed in rust proof compact and sleek cabinet movable on four castors.</p> <p>11. Should be from reputed manufacturer with experience in hemodialysis treatment.</p> <p>12. Should be supplied with power cord and water hoses with quick connectors for machine.</p> <p>13. Should be supplied with bicarbonate mixer.</p> <p>2. Operating Environment</p> <p>2.1 Power supply: 220-230VAC, single phase, 60Hz fitted with appropriate sizing of wire and circuit breaker.</p> <p>3. Standards & Safety Requirements</p> <p>3.1 Shall meet IEC-60601 standard.</p>							

**Page 3 of Request for Quotation for SUPPLY AND DELIVERY OF MEDICAL AND LABORATORY EQUIPMENT
FOR USE OF GEAMH**

No	ITEM DESCRIPTION (ITEM NAME&TECHNICALSPECIFICATIONS)	QTY.	UNIT	APPROVED BUDGET OF CONTRACT (ABC)	FINANCIAL PROPOSAL <i>(Indicate the Price Offer)</i>		TECHNICAL PROPOSAL	Delivery weeks /days
					Unit Price	Total Price		
	<p>4. User Training 4.1 The Supplier shall conduct user and in-house biomedical equipment technician training for this equipment to enable operators and maintenance personnel to use the equipment properly. The training shall include the use of all operational functions of the equipment, as well as routine checks and maintenance expected by users and biomedical equipment technician. 4.2 One week on-site training by manufacturers' application specialist and service engineer for training of medical and technical staffs to familiarize in proper operation and basic maintenance and troubleshooting.</p> <p>5. Warranty 5.1 The warranty is one (1) year after acceptance.</p> <p>6. Maintenance Service During Warranty Period 6.1 During warranty period supplier must ensure preventive maintenance & corrective/breakdown maintenance whenever required. 6.2 Manufacturer shall commit the availability of spare parts and consumables and accessories for at least 5 years.</p> <p>7. Installation and Commissioning 7.1 The bidder must arrange for the equipment to be installed by certified or qualified personnel; any prerequisites for installation to be communicated to the purchaser in advance, in detail.</p> <p>8. Documentation 8.1 Two sets user (Operating) manual in English and soft copy in pdf. 8.2 Two sets service (Technical / Maintenance) manual in English and soft copy in pdf. 8.3 List of important spare parts and accessories with their part numbers and costing. 8.4 Certificate of performance and inspection from factory.</p>							
2.	BICARB MIXER	1	UNIT	85,000.00				
	nothing follows							
	TOTAL			570,000.00				

Company Name.

Address

Print Name & Signature of Authorized Representative